

Ultimate Wireforms, Inc.  
200 Central Street  
Bristol  
Connecticut  
06010  
USA

31 August 2023

**Notified Body Confirmation Letter**  
**Reference: EU2023-607/679110 rev.0**

To whom it may concern,

**Confirmation of the status of a formal application, written agreement, and appropriate surveillance in the framework of Regulation (EU) 2023/607 amending Regulations (EU) 2017/745 and (EU) 2017/746 as regards the transitional provisions for certain medical devices and in vitro diagnostic medical devices**

This letter confirms that, **BSI Group The Netherlands B.V.**, a Notified Body (NB) designated against Regulation (EU) 2017/745 (MDR) and identified by the number **2797** on NANDO, has received a formal application in accordance with Section 4.3, first subparagraph of Annex VII of MDR and has signed a written agreement in accordance with Section 4.3, second subparagraph of Annex VII of MDR with the following manufacturer:

Ultimate Wireforms, Inc.  
200 Central Street  
Bristol  
Connecticut  
06010  
USA  
SRN Number: US-MF-000013501


The devices covered by the formal application and the written agreement mentioned above are identified in the Tables below. Table 1 identifies the devices for which an MDR application has been received, written agreement concluded and for which the NB is also responsible for appropriate surveillance of the corresponding devices under the applicable Directive. Table 2 identifies the devices for which an MDR application has been received and a written agreement concluded, but the NB has not yet taken the responsibility for appropriate surveillance of the corresponding devices under the applicable Directive.

In the case of devices covered by certificates issued under Directive 90/385/EEC (AIMDD) or Directive 93/42/EEC (MDD) that expired after 26 May 2021 and before 20 March 2023, without having been withdrawn, this letter also confirms that the manufacturer signed the written agreement under MDR by the date of MDD/AIMDD certificate expiry; or provided evidence that a competent authority of a Member State had granted a derogation or exemption from the applicable conformity assessment procedure in accordance with Article 59(1) of MDR or Article 97(1) of the MDR respectively, by the 20 Mar 2023 for the relevant devices.

The transition timelines that apply to the devices covered by this letter, subject to the manufacturer's continued compliance to the other conditions specified in Article 120.3c of MDR (as amended by (EU) 2023/607), are shown below:

- 26 May 2026 for Class III custom-made implantable devices
- 31 December 2027 for Class III devices and Class IIb implantable devices excluding Well-established technologies (WET - sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips and connectors)
- 31 December 2028 for other Class IIb devices, Class IIa, Class I devices placed on the market in sterile condition or have a measuring function
- 31 December 2028 for devices not requiring the involvement of a notified body under MDD but requiring it under MDR (e.g., class I devices that qualify as re-usable surgical instruments)

On behalf of BSI Group The Netherlands B.V.,



Digitally signed by  
Dörte Asmus  
Date: 2023.08.31  
11:35:34 +02'00'

Dörte Asmus  
BSI Scheme Manager

**Table 1: Devices covered by this letter and for which the NB is also responsible for appropriate surveillance of the corresponding devices under the applicable Directive:**

<b>Device name or Basic UDI-DI (under MDR application)</b>	<b>MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)</b>	<b>If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device</b>	<b>MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification</b>
NiTi Orthodontic Springs	Class IIa	N/A	MDD Annex II section 3.2 CE 613829 Exp date : 17 SEP 2023 NB# : 2797
Stainless Steel Orthodontic Springs	Class IIa	N/A	MDD Annex II section 3.2 CE 613829 Exp date : 17 SEP 2023 NB# : 2797
Beta III CNA Orthodontic Wires	Class IIa	N/A	MDD Annex II section 3.2 CE 613829 Exp date : 17 SEP 2023 NB# : 2797
NiTi Orthodontic Wires	Class IIa	N/A	MDD Annex II section 3.2 CE 613829 Exp date : 17 SEP 2023 NB# : 2797
Stainless Steel Orthodontic Wires	Class IIa	N/A	MDD Annex II section 3.2 CE 613829 Exp date : 17 SEP 2023 NB# : 2797
Stainless Steel Accessories - Orthodontic Ligature Ties and Kobayashi Ties	Class IIa	N/A	MDD Annex II section 3.2 CE 613829 Exp date : 17 SEP 2023 NB# : 2797
Stainless Steel Accessories (Orthodontic Crimpable Stops & Crimpable Tubes)	Class IIa	N/A	MDD Annex II section 3.2 CE 613829 Exp date : 17 SEP 2023 NB# : 2797

**Table 2: Devices covered by this letter and for which the NB is NOT responsible for appropriate surveillance of the corresponding devices under the applicable Directive:**

Device name or Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
N/A	Choose an item.	N/A	N/A

**Confirmation Letter Revision History**

Date	Action
2023/08/25	Initial issue